

Underbody blankets have a higher heating effect than overbody blankets in lithotomy position endoscopic surgery under general anesthesia: a randomized trial

Running head: Under- and overbody blanket heating

Kentaro Hara, MSN ^{1,2}, Hiromi Kuroda, PhD ², Emi Matsuura, PhD ², Yuji Ishimatsu, PhD ², Sumihisa Honda, PhD ², Hiroaki Takeshita, PhD ³, Terumitsu Sawai, PhD ²

¹ Department of Operation Center, National Hospital Organization Nagasaki Medical Center, Nagasaki, Japan 856-8562

² Nagasaki University Graduate School of Biomedical Sciences, Nagasaki, Japan 852-8523

³ Department of Surgery, National Hospital Organization Nagasaki Medical Center, Nagasaki, Japan 856-8562

Corresponding author: Kentaro Hara

Full preferred mailing address: Kubara 2-1001-1, Omura, Nagasaki 856-8562, Japan

Contact telephone: +81-957-52-3121

Fax number: +81-957-54-0292

Contact e-mail: hara.kentaro.yu@mail.hosp.go.jp

Funding: no financial support was used for this work.

Abstract

Background: Surgery under general anesthesia results in temperature decrease due to the effect of anesthetics and peripheral vasodilation on thermoregulatory centers.

Perioperative temperature control is therefore an issue of high importance. In this study, we aimed to compare the warming effect of underbody and overbody blankets in patients undergoing surgery in the lithotomy position under general anesthesia.

Methods: From September 2018 to October 2019, 99 patients undergoing surgery for colorectal cancer in the lithotomy position were included in this randomized controlled trial and assigned to the intervention group (underbody blanket) or control group (overbody blanket).

Results: The central temperature was significantly higher in the underbody blanket group than in the overbody blanket group at 90 min after the beginning of the surgery ($p = 0.02$); also in this group, the peripheral temperature was significantly higher 60 min after the beginning of the surgery ($p = 0.02$). Regarding postoperative factors, the underbody blanket group had a significantly lower frequency of postoperative shivering ($p < 0.01$) and a significantly shorter postoperative hospital stay ($p = 0.04$) than the overbody blanket group.

Conclusions: We recommend the use of underbody blankets for intraoperative

temperature control in patients undergoing surgery in the lithotomy position under general anesthesia. Underbody blankets showed improved rise and maintenance of central and peripheral temperature, decreased the incidence of postoperative shivering, and shortened the postoperative length of hospital stay.

Keywords: Underbody blanket; Overbody blanket; Lithotomy position; General anesthesia

Introduction

Surgery under general anesthesia causes suppression of the thermoregulatory center and peripheral vasodilation, leading to a decrease in central temperature of 1–3°C [1-3]. It has been reported that intraoperative hypothermia increases surgical site infection (SSI), prolongs the hospital stay, increases the cost of medical care, and increases bleeding risk. The need for maintaining normothermia in the perioperative period has been shown to play a role in postoperative recovery [4-7]. In addition, studies on intraoperative warming and prevention of perioperative complications have reported that hypothermia leads to delayed arousal due to decreased hepatic drug-metabolizing enzyme activity [8-10].

There are four types of heat transfer in the human body: radiation, convection, evaporation, and conduction [11]. Because 90% of the heat is delivered through the skin, this surface is required to safely transfer large amounts of heat to the patient [12]. Specific intraoperative patient warming methods have been used in the past to provide heat through radiation and conduction, and through hot water circulation systems. In recent years, the effectiveness of hot air heaters, which envelop the patient with hot air and transfer heat by convection, contrary to hot water circulation heaters, which only heat the area in contact with the patient, has been reported [13-14].

There are two types of hot air heaters: the overbody blanket, which warms the patient from above, and the underbody blanket, which warms the patient from below. However, with the recent development of laparoscopic surgery, the number of surgeries performed in the lithotomy position has rapidly increased, while the number of cases of intraoperative hypothermia has dramatically increased. Therefore, it is necessary to improve the temperature control methods and the body heating area because lithotomy often involves upper extremity restraining. However, the area of warmth is frequently limited to the neck and part of the anterior thorax. If an underbody blanket is used, the patient is heated from the head to the buttocks, which is likely to be more effective in preventing intraoperative hypothermia. Nevertheless, due to the higher cost of the underbody heating system, it is difficult to implement its widespread use without supporting evidence. Although there have already been previous studies showing the effectiveness of underbody heating systems for supine surgery [15-17], there are no reports on their effectiveness for procedures performed in the lithotomy position.

We considered that it is necessary to carry out basic research to clarify the effect of the underbody blanket in surgeries performed with the patient in the lithotomy position.

Materials and methods

Study setting and population

The study was a single-blind randomized controlled trial conducted between September 2018 and November 2019 in the operating rooms of the National Hospital Organization in Nagasaki, Japan. All participants were at least 20 years old at the time of obtaining consent to perform colon cancer laparoscopic surgery under general anesthesia. Patients were fully informed regarding their participation in the study and confirmed their full understanding before providing written consent, which was obtained after confirming that participation in the study was secure. Exclusion criteria were emergency surgery for colorectal cancer, a family history of malignant high fever, and drug sensitivity (i.e., patients with allergies). The research process was fully explained to the participants and their families. The required number of patients was calculated using G*power 3.1.9.4 (Heinrich-Heine Universität, Dusseldorf, Germany) considering an effect size, power, and α of 0.6, 0.9, and 0.05, respectively. Ninety patients (45 patients in each group) were deemed necessary. A total of 100 patients (50 patients in each group) were enrolled considering a dropout rate of 10%.

Ethical considerations

This study was approved by the Ethics Committee of the Nagasaki University Graduate School of Medicine, Dentistry and Pharmaceutical Sciences (No. 18071204) and by the Ethics Committee of the Nagasaki Medical Center (No. 30071), where the research was conducted. This study was registered with the University Hospital Medical Information Network Center (ID: 000034019). The CONSORT checklist was used to report the study [18].

Design and randomization

We stratified patients into those under and over 70 years old, respectively, taking into account the median age of patients who underwent elective colorectal cancer surgery under general anesthesia between 2016 and 2018. Equal randomization to the intervention (the underbody blanket) and control (the overbody blanket) groups was achieved via a computer-generated random sequence list. Concealed allocation was ensured as the research assistant conducting randomization received the group assignment without access to participant information.

Outcomes

The primary endpoint was intraoperative central temperature. Secondary

endpoints were intraoperative peripheral temperature, intraoperative blood loss, postoperative shivering, postoperative complications, and postoperative hospital length of stay.

As central temperature, we measured the tympanic temperature when patients entered and exited the operating room, and measured the esophageal temperature from the time of induction of anesthesia until the end of surgery: during anesthesia initiation; at the initiation of surgery; 30, 60, 90, 120, and 180 min later; and after surgery completion. The measurement of tympanic temperature shows errors depending on the angle; thus, the average of three tympanic membrane temperature measurements was used to correct for errors due to the angle at the time of measurement.

We measured the skin temperature on the dorsum of the hands at the same timepoints used for the central temperature.

Warming methods

A 3M™ Bear Hugger™ Patient Warming Model 675 (Arizant Healthcare Inc., Prairie, MN) was used as an air heater. The operating table was warmed using the 3M™ Bear Hugger™ Postoperative Blanket Model 300 (Arizant Healthcare Inc., Prairie, MN). The temperature was increased to 43°C 30 min before the patient was scheduled to enter

the operating room.

Preoperative warming from entry into the operating room to the initiation of surgery was performed using the 3M™ Bear Hugger™ postoperative Blanket Model 300(Arizant Healthcare Inc., Prairie, MN). The subjects were warmed at 43°C in the supine position from the time they were lying on the operating table until just before the initiation of surgical positioning.

The heating fluid used was Physio® 140 (Otsuka Pharmaceutical Co., Ltd, Japan), which was warmed in an operating room warming cabinet (set at 37°C).

Intervention group (Underbody blanket group)

A 3M™ Bear Hugger™ Underbody Blanket 585 (Arizant Healthcare Inc., Prairie, MN) was used in the intervention group. To enhance the heating effect, the attached head drape was used on the head and neck of the study participants. To heat the thighs, the strips on either side of the lower body blanket were tied to the thighs of the participants. The hot air heater was activated after the surgical position was fixed and before the start of the surgery so that the hot air was not cut off during surgery.

Control group (Overbody blanket group)

A 3M™ Bear Hugger™ Patient Warming Blanket 622 (Arizant Healthcare Inc., Prairie, MN) was used. The head drapes attached to the head and neck of the participants were used to increase heat retention. Both upper and lower extremities were covered with towelettes for insulation.

Warming area

The warming area of the under- and overbody blankets were calculated using the formula of Lund and Browder [19]. This formula, which is used to calculate the area of burns, was adopted because it is possible to calculate the area of each body part in detail, and it was deemed applicable to the measurement of the heated area in this study.

Anesthesia procedure

After entering the operating room, the electrocardiogram leads, blood pressure cuff, and pulse oximeter were fixed with the patient in the supine position. The anesthesiologist in charge administered 0.6–1.0 mg/kg of propofol, 0.3 µg/kg/min of remifentanyl, and 0.6 mg/kg of rocuronium to induce anesthesia. When sufficient muscle relaxation was obtained, the tracheal tube was intubated into the main bronchus by direct or video laryngoscopy. Anesthesia was maintained using sevoflurane, desflurane,

remifentanyl, or rocuronium in most cases at the discretion of the anesthesiologist in charge. Total intravenous anesthesia with propofol was performed in a few cases. Intraoperative ventilator settings were set at a tidal volume of 6–8 ml/kg, positive end expiratory pressure of 4–8 cm H₂O, respiratory rate of 10–14 cycles/min, and fraction of inspired oxygen of 0.4. An open arterial pressure line was placed in the radial artery of the study subjects. The anesthesiologist administered ephedrine, phenylephrine, dopamine, and hydroxyethyl starch 6% so that the mean blood pressure was greater than 60 mmHg.

Fifteen mg/kg of acetaminophen and 50 mg of nonsteroidal anti-inflammatory drugs were administered for analgesia purposes before the end of surgery. Using a pharmacokinetic simulation, fentanyl was administered so that the effective site concentration at the time of awakening was 1–2 ng/ml. Four mg/kg sugammadex were administered before extubation, which was performed after the stabilization of spontaneous ventilation and the recovery from muscle relaxation. After extubation, the patient was transferred to the postoperative observation room or the intensive care unit.

Data collection and statistical analyses

The primary and secondary endpoints, patient background data, intra- and

postoperative factors, and pre-post surgery differences in median central and peripheral were compared between groups. Two-sided p -values were used, and values under 0.05 were considered to indicate statistical significance. All statistical analyses were performed using JMP[®] 14 (SAS Institute Inc., Cary, NC). A t-test or one-way analysis of variance was used for between-group comparisons of continuous variables if data were normally distributed or Mann-Whitney U test if normality was not found. Chi-squared and Fisher's exact tests were used for discrete variables.

Results

Overview of participants

Among the 106 patients scheduled for surgery during the study period, 100 patients provided consent and were enrolled and randomly allocated into the study groups. Of these, 99 were included in the analysis because one patient withdrew consent after participating in the study. There were no cases lost to follow-up (Figure 1). Patient background data are shown in Table 1. There was no significant difference in patient background characteristics between groups. [Figure 1 should be included near here]

Comparison of intraoperative factors

Intraoperative factors for the intervention and control groups are shown in Table

2. There was no significant difference in intraoperative factors between groups.

Comparison of changes in central temperature

Changes in the central temperature are summarized in Table 3 and Figure 2.

Ninety minutes postoperatively, central temperature was significantly higher in the intervention group. In the control group a temperature lower than 36°C was observed in two subjects. [Figure 2 should be included near here]

Comparison of changes in peripheral temperature

Changes in peripheral temperature are summarized in Table 4 and Figure 3.

Sixty minutes postoperatively, peripheral temperature was significantly higher in the intervention group. [Figure 3 should be included near here]

Comparison of differences between central and peripheral temperatures

The median temperature difference (median central –median peripheral temperature) was 0.7–1.5°C in the intervention group and 1.2–2.1°C in the control group. The difference between the central and peripheral temperatures was significantly

smaller in the intervention group ($p < 0.01$).

Comparison of postoperative factors

Postoperative factors for the intervention and control groups are shown in Table 5. There were no postoperative complications other than SSI in both groups. The intervention group had a significantly lower frequency of postoperative shivering ($p < 0.01$) [20] and a significantly shorter postoperative length of hospital stay ($p = 0.04$) than the control group.

Discussion

Unplanned periodic hypothermia (UPH) has been reported in multiple studies of anesthetized patients, and the importance of normothermia is stated in clinical guidelines [21-24]. The normal central temperature is 36.8–37.0°C; UPH is defined as a central temperature below 36.0°C [25]. It has been reported that the occurrence of UPH leads to the reduction of hemostatic function, the occurrence of ischemic heart disease, increased risk of pressure ulcers, and patient discomfort [26-30]. It is important to prevent UPH by using intraoperative heating. In this study, the use of an underbody blanket was found to prevent UPH compared with an upper body blanket because UPH

was observed in only two cases in the control group.

We hypothesized that the underbody blanket would provide a wider and more effective heating area than the overbody blanket for surgeries performed in the lithotripsy position. The median value of each body surface area showed that about 3.3 times more warm air was provided by the underbody blanket. Differences in the heated area were also reported in previous studies [31]. When the prevention of hypothermia is considered, it has been reported that the clinically effective temperature difference is 0.2°C or more [32]. In this study, there was a difference of 0.4°C between the intervention and control groups in the median central temperature at the end of surgery. We assume that the increased heated area of the underbody blanket played a role in the changes in central temperature recorded in the intervention and control groups during surgery.

Postoperative shivering increases the risk of ischemic heart disease, delayed wound healing, and postoperative pain by increasing oxygen consumption from 300 to 400% and increasing the cardiac burden [33-36]. Shivering after surgery under general anesthesia is frequently preceded by vasoconstriction of apical and toe arteriovenous shunts, an autonomic thermoregulatory response different to symptoms observed during central hypothermia [37]. Vasoconstriction functions to maintain normal central nervous

system temperature by decreasing skin and muscle blood flow and reducing heat loss, with significant effects on body heat distribution [38-39]. The difference between central and peripheral temperature was significantly different between shivering and non-shivering patients. It was reported that shivering occurs in patients whose peripheral temperature decreased during surgery [40]. In this study, the difference between central and peripheral temperature was significantly lower in the intervention group; this is considered to be one of the factors that limits the generation of postoperative shivering. In addition, it has been reported that increased peripheral temperature linearly decreases the central temperature at which arteriovenous shunt vasoconstriction and shivering occur [41]. According to our findings, the peripheral temperature of the intervention group was significantly higher than that of the control group after 60 min of surgery. The use of the underbody blanket enables warm air heating to the end of the upper extremity, which has been proven to increase peripheral temperature. It is therefore assumed that the underbody blanket decreased the incidence of postoperative shivering by maintaining the peripheral temperature and preventing a decrease in central temperature, thereby reducing the difference between them.

The postoperative length of hospital stay was significantly shorter in the intervention group. We hypothesized that preventing intraoperative hypothermia would

lead to the prevention of SSI and affect the postoperative hospital length of stay.

However, there was no significant difference in SSI between groups. There were also no significant differences in disease staging, nutritional status, or blood loss, which may be related to the postoperative hospital length of stay. Significant differences between groups were only observed in intraoperative central and peripheral temperature, postoperative shivering, and length of stay. More specifically, postoperative shivering is related to increased oxygen consumption immediately after surgery, which lowers the oxygen supply to the wounded tissues. This may have influenced the delay in wound healing. In a previous study, it was suggested that not only SSI, but also the initiation of solid food and the removal of sutures, affected the postoperative hospital length of stay [5]. In this study, we analyzed the relationship between intraoperative central temperature and SSI, although we did not collect any information on the relationship between SSI and initiation of solid food intake or removal of sutures. In future studies, we will aim to clarify the relation between intraoperative central temperature, suture removal, and initiation of solid food, and their effects in the postoperative hospital length of stay.

The results of this study showed that the use of an underbody blanket in surgery performed in the lithotomy position under general anesthesia led to an increase in

central temperature compared with an overbody blanket. It also contributed to prevent postoperative shivering by decreasing the difference between central and peripheral temperature. We, therefore, recommend performing intraoperative temperature control using an underbody blanket in surgeries performed in the lithotomy position under general anesthesia.

Limitations

The results of this study may not be applicable to all surgeries performed with general anesthesia because this was a single-center, single-disease, randomized controlled trial. Moreover, intraoperative heating was stopped in 36% of cases in the intervention group after consulting the anesthesiologist due to persistent hyperthermia. Hyperthermia at the time of awakening may lead to patient discomfort. In addition, it was decided that the timing to stop intraoperative heating should be decided after consultation with the anesthesiologist in charge, since the appropriate intraoperative central temperature has not been clearly determined. In the future, the range of intraoperative normothermic central temperature to be maintained even after the warm air heater is stopped should be determined.

In conclusion, the use of the underbody blanket had the following effects in

comparison with the overbody blanket in the lithotomy position surgery under the general anesthesia. Adequate intraoperative central and peripheral temperature maintenance appears to have a preventing effect on postoperative shivering and to shorten the postoperative hospital length of stay.

Acknowledgements

We gratefully acknowledge the work of past and present members of our medical center. We would like to thank Editage (www.editage.com) for English language editing.

Disclosure

Kentaro Hara, MSN, Hiromi Kuroda, PhD, Emi Matsuura, PhD, Yuji Ishimatsu, PhD, Sumihisa Honda, PhD, Hiroaki Takeshita, PhD, and Terumitsu Sawai, PhD have no conflicts of interest or financial ties to disclose.

References

1. Morris RH, Wilkey BR (1970) The effects of ambient temperature on patient temperature during surgery not involving body cavities. *Anesthesiology* 32:102-107
2. Morris RH (1971) Influence of ambient temperature on patient temperature during intraabdominal surgery. *Ann Surg* 173:230-233
3. Morris RH (1971) Operating room temperature and the anesthetized, paralyzed patient. *Arch Surgery* 102:95-97
4. Kurz A, Sessler DI, Lenhardt R (1996) Perioperative normothermia to reduce the incidence of surgical-wound infection and shorten hospitalization. *N Engl J Med* 334:1209-1215
5. Gustafsson UO, Scott MJ, Schwenk W, Demartines N, Roulin D, Francis N, McNaught CE, Macfie J, Liberman AS, Soop M, Hill A, Kennedy RH, Lobo DN, Fearon K, Ljungqvist O; Enhanced Recovery After Surgery (ERAS) Society, for Perioperative Care; European Society for Clinical Nutrition and Metabolism (ESPEN); International Association for Surgical Metabolism and Nutrition (IASMEN) (2013) Guidelines for perioperative care in elective colonic surgery: Enhanced Recovery After Surgery (ERAS[®]) Society recommendations. *World J Surg* 37:259-284
6. Mason SE, Kinross JM, Hendricks J, Arulampalam TH (2017) Postoperative

hypothermia and surgical site infection following peritoneal insufflation with warm, humidified carbon dioxide during laparoscopic colorectal surgery: a cohort study with cost-effectiveness analysis. *Surg Endosc* 31:1923-1929

7. Mason SE, Kinross JM, Hendricks J, Arulampalam TH (2017) Postoperative hypothermia and surgical site infection following peritoneal insufflation with warm, humidified carbon dioxide during laparoscopic colorectal surgery: a cohort study with cost-effectiveness analysis. *Surg Endosc* 31:1923-1929

8. Nieh HC, Su SF (2016) Meta-analysis: effectiveness of forced-air warming for prevention of perioperative hypothermia in surgical patients. *J Adv Nurs* 72:2294-2314

9. Ousey K, Edward KL, Lui S, Stephenson J, Walker K, Duff J, Leaper D (2017) Perioperative, local and systemic warming in surgical site infection: a systematic review and meta-analysis. *J Wound Care* 26:614-624

10. Su SF, Nieh HC (2018) Efficacy of forced-air warming for preventing perioperative hypothermia and related complications in patients undergoing laparoscopic surgery: a randomized controlled trial. *Int J Nurs Pract* 24:e12660

11. English MJ, Farmer C, Scott WA (1990) Heat loss in exposed volunteers. *J Trauma* 30:422-425

12. Bräuer A, Quintel M (2009) Forced-air warming: technology, physical background

and practical aspects. *Curr Opin Anesthesiol* 22:769-774

13. Sessler DI (1997) Mild perioperative hypothermia. *N Engl J Med* 336:1730-1737

14. Luck AJ, Moyes D, Maddern GJ, Hewett PJ (1999) Core temperature changes during open and laparoscopic colorectal surgery. *Surg Endosc* 13:480-483.

15. Insler SR, Bakri MH, Nageeb F, Mascha E, Mihaljevic T, Sessler DI (2008) An evaluation of a full-access underbody forced-air warming system during near-normothermic, on-pump cardiac surgery. *Anesth Analg* 106:746-750

16. Pu Y, Cen G, Sun J, Gong J, Zhang Y, Zhang M, Wu X, Zhang J, Qiu Z, Fang F (2014) Warming with an underbody warming system reduces intraoperative hypothermia in patients undergoing laparoscopic gastrointestinal surgery: a randomized controlled study. *Int J Nurs Stud* 51:181-189

17. Engelen S, Himpe D, Borms S, Berghmans J, Van Cauwelaert P, Dalton JE, Sessler DI (2011) An evaluation of underbody forced-air and resistive heating during hypothermic, on-pump cardiac surgery. *Anaesthesia* 66:104-110

18. Schulz KF, Altman DG, Moher D, CONSORT Group (2010) CONSORT 2010 statement: updated guidelines for reporting parallel group randomized trials. *BMJ* 340:c332

19. Lund CC (1944) The estimation of areas of burns. *Surg Gynecol Obstet* 79:352-358

20. Badjatia N, Strongilis E, Gordon E, Prescutti M, Fernandez L, Fernandez A, Buitrago M, Schmidt JM, Ostapkovich ND, Mayer SA (2008) Metabolic impact of shivering during therapeutic temperature modulation: the Bedside Shivering Assessment Scale. *Stroke* 39:3242-3247
21. Wagner VD (2010) Patient safety chiller: unplanned perioperative hypothermia. *AORN J* 92:567-571
22. Lynch S, Dixon J, Leary D (2010) Reducing the risk of unplanned perioperative hypothermia. *AORN J* 92:553-565
23. AORN Recommended Practices Committee (2007) Recommended practices for the prevention of unplanned perioperative hypothermia. *AORN J* 85:972-988
24. Torossian A, TEMMP (Thermoregulation in Europe Monitoring and Managing Patient Temperature) Study Group (2007) Survey on intraoperative temperature management in Europe. *Eur J Anesthesiol* 24:668-675
25. Forbes SS, Eskicioglu C, Nathens AB, Fenech DS, Laflamme C, McLean RF, McLeod RS, Best Practice in General Surgery Committee, University of Toronto (2009) Evidence-based guidelines for prevention of perioperative hypothermia. *J Am Coll Surg* 209:492-503
26. Michelson AD, MacGregor H, Barnard MR, Kestin AS, Rohrer MJ, Valeri CR

(1994) Reversible inhibition of human platelet activation by hypothermia in vivo and in vitro. *Thromb Haemost* 71:633-640

27. Frank SM, Fleisher LA, Breslow MJ, Higgins MS, Olson KF, Kelly S, Beattie C

(1997) Perioperative maintenance of normothermia reduces the incidence of morbid cardiac events: a randomized clinical trial. *JAMA* 277:1127-1134

28. Scott EM, Buckland R (2006) A systematic review of intraoperative warming to prevent postoperative complications. *AORN J* 83:1090-1113

29. Sheffield CW, Sessler DI, Hopf HW, Schroeder M, Moayeri A, Hunt TK, West JM

(1996) Centrally and locally mediated thermoregulatory responses alter subcutaneous oxygen tension. *Wound Repair Regen* 4:339-345

30. Sessler DI (2008) Temperature monitoring and perioperative thermoregulation.

Anesthesiology 109:318-338

31. Miyazaki R, Hoka S (2019) What is the predictor of the intraoperative body

temperature in abdominal surgery? *J Anesth* 33:67-73

32. National Collaborating Centre for Nursing and Supportive Care (UK) (2008) The

management of inadvertent perioperative hypothermia in adults. Royal College of

Nursing (UK), London

33. Pauca AL, Savage RT, Simpson S, Roy RC (1984) Effect of pethidine, fentanyl, and

- morphine on postoperative shivering in man. *Acta Anaesthesiol Scand* 28:138-143
34. Sessler DI, Israel D, Pozos RS, Pozos M, Rubinstein EH (1988) Spontaneous postanesthetic tremor does not resemble thermoregulatory shivering. *Anesthesiology* 68:843-850
35. Pflug AE, Aasheim GM, Foster C, Martin RW (1978) Prevention of postanesthesia shivering. *Can Anaesth Soc J* 25:43-49
36. Sharkey A, Gulden RH, Lipton JM, Giesecke AH (1993) Effect of radiant heat on the metabolic cost of postoperative shivering. *Br J Anaesth* 70:449-450
37. Sessler DI, Rubinstein EH, Moayeri A (1991) Physiologic responses to mild perianesthetic hypothermia in humans. *Anesthesiology* 75:594-610
38. Sessler DI, McGuire J, Moayeri A, Hynson J (1991) Isoflurane-induced vasodilation minimally increases cutaneous heat loss. *Anesthesiology* 74:226-232
39. Hynson JM, Sessler DI, Moayeri A, McGuire J, Schroeder M (1993) The effects of preinduction warming on temperature and blood pressure during propofol/nitrous oxide anesthesia. *Anesthesiology* 79:219-228
40. Kuroki C, Godai K, Hasegawa-Moriyama M, Kuniyoshi T, Matsunaga A, Kanmura Y, Kuwaki T (2014) Perfusion index as a possible predictor for postanesthetic shivering. *J Anesth* 28:19-25

41. Cheng C, Matsukawa T, Sessler DI, Ozaki M, Kurz A, Merrifield B, Lin H, Olofsson P (1995) Increasing mean skin temperature linearly reduces the core-temperature thresholds for vasoconstriction and shivering in humans. *Anesthesiology* 82:1160-1168

Figure legends




Figure 1. Trial CONSORT diagram. Among the 106 patients scheduled for surgery during the study period, 100 patients provided consent and were enrolled and randomly allocated into the study groups. Of these, 99 were included in the analysis because one patient withdrew consent after participating in the study.




Figure 2. Ninety minutes postoperatively, central temperature was significantly higher in the intervention group.




Figure 3. Sixty minutes postoperatively, peripheral temperature was significantly higher in the intervention group.

Tables

Table 1. Patient characteristics

| | Underbody blanket | Overbody blanket | <i>p</i>-value |
|--------------------------------------|--------------------------|-------------------------|-----------------------|
| | (N = 50) | (N = 49) | e |
| Age (years) | 71 (35–87) | 70 (24–91) | 0.68 |
| Sex | | | 0.18 |
| Male | 23 (46%) | 29 (59.2%) | |
| Female | 27 (54%) | 20 (40.8%) | |
| Height (cm) | 158.3 (141.6–176) | 159.9 (133–177.7) | 0.98 |
| Weight (kg) | 56 (35.8–81.8) | 57 (35–80.1) | 0.96 |
| Body surface area (cm ²) | 15639 (12108–18529) | 16002 (11630–18924) | 0.98 |
| Body mass index (kg/m ²) | 22.3 (16.6–29.1) | 21.9 (18.1–29.7) | 0.71 |
| Preoperative total protein (g/dl) | 6.9 (5–7.9) | 6.7 (4.7–9.9) | 0.10 |
| Preoperative albumin (g/dl) | 3.9 (2.2–4.7) | 3.8 (2.2–5) | 0.37 |
| Preoperative hemoglobin (g/dl) | 12 (7.7–15.2) | 11.6 (6.7–20.1) | 0.65 |
| Stages of cancer | | | 0.68 |
| Stage I | 8 (16%) | 12 (24.5%) | |
| Stage II | 10 (20%) | 11 (22.4%) | |
| Stage III | 24 (48%) | 20 (40.8%) | |
| Stage IV | 8 (16%) | 6 (12.2%) | |

Note. Values are presented as medians (ranges) or number of patients (%).

Mann–Whitney–U test was used for comparisons of ordinal data. Nominal data were compared using the Chi-squared test.

Table 2. Intraoperative factors

| | Underbody blanket (N = 50) | Overbody blanket (N = 49) | <i>p</i>-value |
|--------------------------------------|---|--------------------------------------|-----------------------|
| Preoperative room temperature (°C) | 26 (25–27) | 26 (25–27) | 0.54 |
| Intraoperative room temperature (°C) | 23.5 (22.5–24.5) | 23.5 (22–24.5) | 0.22 |
| Postoperative room temperature (°C) | 26 (24.5–27) | 26 (24.5–27) | 0.84 |
| Operative time (min) | 324 (178–900) | 330 (182–671) | 0.98 |
| Anesthetic time (min) | 414.5 (252–977) | 453 (225–816) | 0.81 |
| Type of anesthesia | | | 0.57 |
| Sevoflurane | 29 (58%) | 32 (65.3%) | |
| Desflurane | 19 (38%) | 14 (28.6%) | |
| Propofol | 2 (4%) | 3 (6.1%) | |
| Amount of bleeding (ml) | 30 (2–795) | 30 (3–635) | 0.75 |
| Urine volume (ml) | 332.5 (70–2250) | 30 (3–635) | 0.64 |
| Total fluid volume (ml) | 2233 (1240–4819) | 2553 (1039–5719) | 0.15 |
| Warming area (cm ²) | 5708 (4419–6763) | 2400 (1744–2839) | < 0.01 * |
| Infusion warming apparatus | | | 0.11 |
| Used | 0 (0%) | 3 (6.1%) | |

| | | | |
|--|-----------|------------|------|
| Not used | 50 (100%) | 46 (93.9%) | |
| Intraoperative blood transfusion | | | 0.61 |
| Required | 1 (2%) | 2 (4.1%) | |
| Not required | 49 (98%) | 47 (95.9%) | |
| Intraoperative warming blanket stopped | | | 0.61 |
| Stopped | 18 (36%) | 10 (20.4%) | |
| Not stopped | 32 (64%) | 39 (79.6%) | |

Note. Values are presented as medians (ranges) or number of patients (%).

Mann–Whitney–U test was used for comparisons of ordinal data. Nominal data were compared using the Chi-squared test or Fisher's exact test.

*significant difference between groups

Table 3. Between-group differences in intraoperative central temperature over time

| | Underbody blanket (N = 50) | Overbody blanket (N = 49) | p-value |
|---------------------------------------|---|--------------------------------------|----------------|
| When entering the operation room (°C) | 36.3 (35.1–37.5) | 36.4 (34.9–37.7) | 0.51 |
| Anesthesia initiation(°C) | 36.5 (35.1–37.3) | 36.4 (35.3–37.6) | 0.52 |
| Initiation of surgery (°C) | 36.5 (35.6–37.2) | 36.4 (35.5–37.6) | 0.48 |
| 30 min later (°C) | 36.7 (35.5–37.2) | 36.6 (35.4–37.5) | 0.41 |
| 60 min later (°C) | 36.9 (35.6–37.2) | 36.6 (35.7–37.5) | 0.09 |
| 90 min later (°C) | 37.0 (35.8–37.7) | 36.7 (35.8–37.7) | 0.02* |
| 120 min later (°C) | 37.1 (35.6–37.8) | 36.8 (35.7–37.9) | 0.01* |
| 180 min later (°C) | 37.2 (36.1–37.9) | 36.8 (35.7–37.9) | 0.03* |
| Completion of surgery (°C) | 37.5 (36.4–38.6) | 37.1 (35.2–38.2) | < 0.01 * |
| When leaving the operating room (°C) | 37.0 (35.8–38.5) | 36.7 (35.3–38.3) | 0.05 |

Note. Values are presented as medians (ranges) or number of patients (%).

Mann–Whitney–U test was used for comparisons of ordinal data.

*significant difference between groups

Table 4. Between-group differences in intraoperative peripheral temperature over

| | time | | |
|---------------------------------------|---|--------------------------------------|----------------|
| | Underbody blanket (N = 50) | Overbody blanket (N = 49) | p-value |
| When entering the operation room (°C) | 32.0 (27.4–36.6) | 32.3 (26.9–37.0) | 0.18 |
| Anesthesia initiation (°C) | 33.0 (27.5–36.2) | 33.0 (28.4–36.8) | 0.59 |
| Initiation of surgery (°C) | 35.0 (31.1–38.4) | 35.0 (30.6–36.6) | 0.96 |
| 30 min later (°C) | 35.7 (31.5–38.0) | 35.0 (32.0–36.9) | 0.39 |
| 60 min later (°C) | 36.0 (32.2–38.4) | 35.3 (32.5–36.9) | 0.02* |
| 90 min later (°C) | 36.2 (32.0–38.5) | 35.5 (32.0–36.9) | < 0.01 * |
| 120 min later (°C) | 36.2 (32.0–38.6) | 35.5 (32.0–37.1) | < 0.01 * |
| 180 min later (°C) | 36.4 (31.1–38.3) | 35.2 (32.4–37.3) | < 0.01 * |
| Completion of surgery (°C) | 36.1 (29.8–38.8) | 35.0 (29.4–37.3) | 0.03* |
| When leaving the operating room (°C) | 35.7 (29.4–37.7) | 35.0 (26.6–36.9) | 0.22 |

Note. Values are presented as medians (ranges) or number of patients (%).

Mann–Whitney–U test was used for comparisons of ordinal data.

*significant difference between groups

Table 5. Postoperative factors

| | Underbody blanket (N = 50) | Overbody blanket (N = 49) | p-value |
|---|---|--------------------------------------|----------------|
| Postoperative shivering | | | < 0.01 * |
| Yes | 1 (2%) | 9 (18.4%) | |
| Score 1 | 1 (100%) | 2 (22.2%) | |
| Score 2 | 0 | 6 (66.7%) | |
| Score 3 | 0 | 1 (11.1%) | |
| No | 49 (98%) | 40 (81.6%) | |
| Surgical site infection | | | 0.36 |
| Yes | 4 (8%) | 1 (2%) | |
| No | 46 (92%) | 48 (98%) | |
| Postoperative hospitalization (days) | 11 (6–75) | 13 (7–42) | 0.04* |

Note. Values are presented as medians (ranges) or number of patients (%).

Mann–Whitney–U test was used for comparisons of ordinal data. Nominal data were compared using the Fisher's exact test.

*significant difference between groups